

Title	Challenges in the interpretation of colorectal indocyanine green fluorescence angiography: Video vignette
Authors	Dalli, Jeffrey;Hardy, Niall;Mac Aonghusa, Pol G.;Epperlein, Jonathan P.;Cantillon-Murphy, Pádraig;Cahill, Ronan A.
Publication date	2021-02-18
Original Citation	Dalli, J.,Hardy, N., Mac Aonghusa, P.G., Epperlein, J. P., Cantillon Murphy, P. and Cahill, R. A. (2021) 'Challenges in the interpretation of colorectal indocyanine green fluorescence angiography: Video vignette', Colorectal Disease, 23(5), pp.1289-1290. https://doi.org/10.1111/codi.15592
Type of publication	Article (peer-reviewed)
Link to publisher's version	10.1111/codi.15592
Rights	© 2021, John Wiley & Sons, Inc. This is the peer reviewed version of the following item: Dalli, J.,Hardy, N., Mac Aonghusa, P.G., Epperlein, J. P., Cantillon Murphy, P. and Cahill, R. A. (2021) 'Challenges in the interpretation of colorectal indocyanine green fluorescence angiography: Video vignette', Colorectal Disease, doi: 10.1111/codi.15592, which has been published in final form at https://doi.org/10.1111/codi.15592 . This article may be used for non-commercial purposes in accordance with Wiley Terms and Conditions for Use of Self-Archived Versions.
Download date	2026-09-24 04:17:10
Item downloaded from	https://hdl.handle.net/10468/11102

DR JEFFREY DALLI (Orcid ID : 0000-0003-0170-3667)

Article type : Correspondence

DR JEFFREY DALLI (Orcid ID: 0000-0003-0170-3667)

Article type: Correspondence

Corresponding Author

Mr Jeffrey Dalli MD MRCSEd MSc FEBS

jeffreydalli@gmail.com

ORCID - 0000-0003-0170-3667

UCD Centre for Precision Surgery,

University College Dublin

Catherine McAuley Centre

21 Nelson St, Phibsborough,

Dublin 7 D07 KX5K, Ireland

Challenges in the interpretation of colorectal indocyanine green fluorescence angiography: Video Vignette

This article has been accepted for publication and undergone full peer review but has not been through the copyediting, typesetting, pagination and proofreading process, which may lead to differences between this version and the [Version of Record](#). Please cite this article as [doi: 10.1111/CODI.15592](https://doi.org/10.1111/CODI.15592)

This article is protected by copyright. All rights reserved

Jeffrey Dalli^{1,2}, Niall Hardy^{1,2}, Pol G Mac Aonghusa³, Jonathan P Epperlein³, Padraig Cantillon Murphy^{4,5},
Ronan A Cahill^{1,2}

¹ UCD Centre for Precision Surgery, School of Medicine, University College Dublin, Ireland,

² Department of Surgery, Mater Misericordiae University Hospital, Dublin, Ireland,

³ IBM Research Europe, Dublin, Ireland, ⁴School of Engineering, University College Cork, Cork, Ireland,

⁵ Tyndall National Institute, Dyke Parade, Cork, Ireland

Colorectal anastomotic leakage remains a serious complication with implications on hospital stay, oncological outcomes(1) and treatment cost. (2) Traditional intra-operative visual perfusion assessment has been shown to be suboptimal(3) and surgeons are looking to indocyanine green (ICG) fluorescence angiography as an adjunct to clinical judgement. A significant body of work demonstrated feasibility of use(4),(5) and a recent randomised trial showed positive results,(6) however not all studies have found similarly.(7) Although clinical trials are ongoing, signal sensing and presentation have not been standardised. Inter-brand variation in image stack and camera performance are further confounded by distance and ICG concentration effects.(8) Even when using the same equipment, inter-user variability especially among inexperienced users has been demonstrated.(9)

In the following video, we describe ICG 'creep' as different portions of the bowel achieve peak intensity over a time continuum. Delaying interrogation and interpretation for example while positioning the stapler, may result in perceived margin migration. This phenomenon is already accounted for in some upper gastro-intestinal surgery recommendations, that adequate perfusion is to be discriminated at ninety seconds.(10)

Secondly, we use temporal-fluorescence signal profiles to direct attention to the timing of these peaks and troughs. (11),(12) Using a bespoke tracker, (13) signal profile interrogation revealed lower absolute intensities at the periphery of the image despite well perfused inflows. Operators are thus prone to overestimate central image perfusion.

Such human inter-user interpretation inconsistencies may result in suboptimal capitalisation of fluorescence angiography. While manufactures emphasise intensity via hybridised images with glowing colours, accurate quantitative fluorescence signal appreciation may in fact only be reliably achieved via computer vision.

What does this paper add to the existing literature?

This video publication illustrates some of the difficulties in the visual interpretation of the fluorescence angiogram in colorectal surgery anastomosis. Surgical users should be aware of these pitfalls however better interpretation may require decision support in the form of artificial intelligence.

Ethics and Patient Statement: Research was carried out in accordance with relevant guidelines, regulations and with institutional (Mater Misericordiae University Hospital, Dublin, Ireland) ethics approval (1/378/2092, ClinicalTrials.gov Identifier: NCT04220242) Informed consent was obtained from all participants.

Author contributions: All the authors contributed to the design, implementation of the research, analysis of the results and to the writing of the manuscript.

Funding/Financial Support: This project is supported by government of the Republic of Ireland via the Disruptive Technologies Innovation Fund.

Disclosure: R.C. is named on a patent filed in relation to processes for visual determination of tissue biology, receives speaker fees from Stryker Corp, research funding from Intuitive Corp and holds research funding from the Irish Government in collaboration with IBM Research in Ireland. J.D. and N.H. are employed as researchers on the same collaboration. J.P.E. and P.M.A. are full-time employees of IBM Research, a division of IBM, which provides technical products and services worldwide to government, healthcare, and life-sciences companies. P.C.M. is a consultant to Quadrant Scientific Limited.

Acknowledgements: We would like to thank Kilian O'Donoghue at Quadrant Scientific Limited, Cork, Ireland for professional technical support.

References

1. Noh GT, Ann YS, Cheong C, Han J, Cho MS, Hur H, et al. Impact of anastomotic leakage on long-term oncologic outcome and its related factors in rectal cancer. *Medicine (Baltimore)*. 2016;95(30):e4367-e.
2. La Regina D, Di Giuseppe M, Lucchelli M, Saporito A, Boni L, Efthymiou C, et al. Financial Impact of Anastomotic Leakage in Colorectal Surgery. *J Gastrointest Surg*. 2019;23(3):580-6.

3. Karliczek A, Harlaar NJ, Zeebregts CJ, Wiggers T, Baas PC, van Dam GM. Surgeons lack predictive accuracy for anastomotic leakage in gastrointestinal surgery. *International Journal of Colorectal Disease*. 2009;24(5):569-76.
4. Ris F, Liot E, Buchs NC, Kraus R, Ismael G, Belfontali V, et al. Multicentre phase II trial of near-infrared imaging in elective colorectal surgery. *Br J Surg*. 2018;105(10):1359-67.
5. Jafari MD, Wexner SD, Martz JE, McLemore EC, Margolin DA, Sherwinter DA, et al. Perfusion assessment in laparoscopic left-sided/anterior resection (PILLAR II): a multi-institutional study. *J Am Coll Surg*. 2015;220(1):82-92 e1.
6. Alekseev M, Rybakov E, Shelygin Y, Chernyshov S, Zarodnyuk I. A study investigating the perfusion of colorectal anastomoses using fluorescence angiography: results of the FLAG randomized trial. *Colorectal Disease*. n/a(n/a).
7. De Nardi P, Elmore U, Maggi G, Maggiore R, Boni L, Cassinotti E, et al. Intraoperative angiography with indocyanine green to assess anastomosis perfusion in patients undergoing laparoscopic colorectal resection: results of a multicenter randomized controlled trial. *Surg Endosc*. 2020;34(1):53-60.
8. Kono Y, Ishizawa T, Tani K, Harada N, Kaneko J, Saiura A, et al. Techniques of Fluorescence Cholangiography During Laparoscopic Cholecystectomy for Better Delineation of the Bile Duct Anatomy. *Medicine (Baltimore)*. 2015;94(25):e1005-e.
9. Hardy NP, Dalli J, Khan MF, Andrejevic P, Neary PM, Cahill RA. Inter-user variation in the interpretation of near infrared perfusion imaging using indocyanine green in colorectal surgery. *Surgical Endoscopy*. 2021.
10. Kumagai Y, Hatano S, Sobajima J, Ishiguro T, Fukuchi M, Ishibashi KI, et al. Indocyanine green fluorescence angiography of the reconstructed gastric tube during esophagectomy: efficacy of the 90-second rule. *Dis Esophagus*. 2018;31(12).
11. Son GM, Kwon MS, Kim Y, Kim J, Kim SH, Lee JW. Quantitative analysis of colon perfusion pattern using indocyanine green (ICG) angiography in laparoscopic colorectal surgery. *Surg Endosc*. 2019;33(5):1640-9.
12. Cahill RA, O'Shea DF, Khan MF, Khokhar HA, Epperlein JP, Mac Aonghusa PG, et al. Artificial intelligence indocyanine green (ICG) perfusion for colorectal cancer intra-operative tissue classification. *British Journal of Surgery*. 2020;108(1):5-9.
13. Zhuk S, Epperlein J, Nair R, Thirupati S, Aonghusa P, Cahill R, et al. Perfusion Quantification from Endoscopic Videos: Learning to Read Tumor Signatures2020.